






$\sqrt{\text{Rz } 1,6} \quad ( \checkmark )$



FOR USE IN CLEAN ENVIRONMENT ISO 7 BY ISO 14644-1  
FOR USE IN HIGH VACUUM  $10^{-6}$  Pa

The informations contained in this drawing are the sole property of FZU. Any reproduction is prohibited				 beamlines	 <b>Fyzikální ústav</b> Akademie věd ČR, v. v. i.
Drawn by: andrlík	Comments		Scale 1:1	Dwg. title BA_Chamber_Rohr 154x2	
Checked by:					
Date: 29.04.2019					
Material: AISI 316 L		All dimensions in mm		Dwg. no. – TC ID/Rev. 00198353/00	
Raw mat.:		Tolerance: ISO2768–mk			
Weight: 1,12 Kg		Application			
ELI B. Nr. <span style="float: right;">Sheet 1 of 1</span>					